

Optical/Electrical Components Needed for iKAGRA IMC Servo

Yuta Michimura


Ando Group

Department of Physics, University of Tokyo


Scope

- Start discussion on listing up what we need for iKAGRA IMC (and PSL) servo
- Focus mainly on optical and electrical components
- References:
 - [JGW-T1402349](#) (iKAGRA PMC study)
 - [JGW-G1402302](#) (FSS modeling)
 - [JGW-G1402520](#) (GWADW2014 IOO poster by Nakano)
 - [JGW-D1402507](#) (IOO 3D drawing)
 - [JGW-T1302068](#) (layout around IMMT)
 - [JGW-D1402492](#) (IMC suspension cabling)
 - [JGW-D1402516](#) (anchor, floor mortar, floor cutting lines)
 - [Wiki/SmallOptics](#) (list of small optics)
 - [Wiki/OutputTables](#) (list of output optical tables)


Optical Configuration




Suspension Local Damping



Initial Alignment




Intensity Stabilization




ISS is not a default plan in iKAGRA


PMC Servo





Frequency Stabilization



Alignment Sensing and Control



Cameras



6 cameras?

How do we put their signals in the digital system?

Gouy Phase Telescopes

- We need two QPDs for each port we want to monitor the beam alignment
- It would be nice to have standardized Gouy phase telescopes for each port
- Below is an example aLIGO one (~ 15 cm x 55 cm)
[LIGO-T1000247](#)





Figure 2: The REFL port QPD layout.

Figure 1: The REFL port telescope beampath.

Optical Components Missing

- PSL periscope
 - we have dumped rods available in Ando Lab
- [Newport 45](#)
[Newport M-340-RC](#) already delivered to Ohshima-san
- Gouy phase telescopes and other optical components for each port
 - we can buy flat mirrors and mounts with a rough estimate of numbers, but what do we do for lenses?
- High voltage amplifier for broadband EOM
[Newport 3211](#) (0.5 MHz bandwidth)
- AOM for FRC?
was already there at ICRR (including a driver)
- I suppose there are much more.....

FSS Topology (rough idea)

- IMC servo (2 MHz/rtHz @ 0.1 Hz, 0.1 Hz/rtHz @ 10 Hz)
 - AOM (common mode board; DC – 100 kHz) AOM: 3110-197
driver: 1110AF-AEFO-1.5
(Crystal Technology)
efficiency: 5.3 MHz/V, range: 90-130 MHz
 - MCE (digital system; DC – 1 Hz; optional)

Actuation efficiency and range?

- FRC servo (with rough estimate, 170 Hz/rtHz @ 1kHz)
 - laser temp (digital system; DC – 0.1 Hz)
efficiency: 3 GHz/V, range: 30 GHz
 - laser PZT (FSS board; DC – 1 kHz)
efficiency: 1 MHz/V, range: 100 MHz
 - EOM (FSS board; 1 kHz – 1 MHz)
efficiency: $0.01 * (f/1 \text{ Hz}) \text{ Hz/V}$, range: 0.6 MHz

Numbers to remember:

- Kamioka seismic noise 2e-7 m/rtHz @ 0.1 Hz, 1e-11 m/rtHz @ 10 Hz, 1e-13 m/rtHz @ 100 Hz
- IMC suspension vibration isolation ratio: 0 dB @ 0.1 Hz, -60 dB @ 10 Hz, -140 dB @ 100 Hz
- IMC round-trip length: $2 * 26.65 \text{ m}$
- laser frequency: $2.8e14 \text{ Hz}$
- NPRO free-run frequency noise: 10 kHz/f Hz/rtHz

DGS Related

- standalone digital system is needed by Nov 2014
- it is not considered as a baseline DGS schedule
 - full digital system with networks will be available only from Jan 2015
- VIS will also need a standalone for IMC suspensions
 - maybe one standalone will do for VIS & IOO

AEL Related

- PMC(1) / FSS(1) / IMC(1) servo circuits
 - IMC servo board is not there yet ([LIGO-D040180](#) aLIGO common mode servo board)
 - TTFSS servo board is already there at ICRR
 - [JGW-D1301823](#) TTFSS board is for EOM path and PZT path (copy of [LIGO-D040105](#))
- RF PD(4) / RF QPD(4) / DC PD(7) / DC QPD(4) (at least!)

PMCREFL IMCREFL REFL FRCREFL	IMCREFL A/B REFL A/B	PMCREFL PMCTrans IMCREFL IMCTrans REFL ISS1 ISS2	IMCTrans A/B ISS1 A/B (it would be nice to have 2 more at PSL table for incident beam monitor)
---------------------------------------	-------------------------	--	--
- IQ demodulators(20ch)
 - 1 for each RF PD, 4 for each RF QPD
- RF distributors
 - f_PMC: split into 2 (EOM, PMCREFL)
 - f_IMC: split into 11 (EOM, IMCREFL, 4xIMCREFLA/B, FRCREFL)
- RF source(3)
 - f_1: split into 21? (EOM, REFL, 4xREFLA/B, AS, 4xASA/B, POX?, POY?)
- picomotor drivers(18ch, at least!)
 - 2 for each mirror
 - PSLSTM1/2
 - STM1/2
 - IMMT1/2
 - MCi/e/o (top stage)
 - (it would be nice to have 2 for each PD/QPD for aligning the beam into them)
- PZT drivers(4ch)
 - 2 for periscope mirror
 - 1 for laser
 - 1 for PMC
- coil drivers(12ch)
 - 4 for each MC mirror

Picomotor Driver


- For CLIO, NewFocus drivers below were used
PICOMOTOR EHTERNET CONTROLLER 8752
INTELLIGENT PICOMOTOR DRIVER 8753
Now, they are discontinued and replaced with
[Four-Axis Picomotor Controller/Driver Kit 8742-4-KIT](#)
- Picomotor control with EPICS and Python
[http://gwwiki.icrr.u-tokyo.ac.jp/JGWwiki/CLIO/Technicals/PMEPICS](#)
[http://gwclio.icrr.u-tokyo.ac.jp/lcgtsubgroup/digitalsystem/2012/03/new-focus-picomotor-controlled-at-stda.html](#)
- No special interface circuit is needed for controlling picomotor from digital system

Cables

- TBD
- [JGW-D1402492](#) (IMC suspension cabling)


Beams around MCE

[JGW-D1402507](#)



Beams around MCI and MCo


[JGW-D1402507](#)



Optical Tables / Pylons Layout

- chamber: ○ pylon: ○ table: □

[Wiki/OutputTables](#)



Rough estimate of space required

oplev input: 0.1x0.1 m

oplev output: 0.1x0.1 m

PD + 2xQPD: 0.8x0.8 m

References:

[JGW-G1402396](#) (oplev)

[LIGO-D1201103](#) (aLIGO ISCT1 table)

We could also place a 1 m x 1 m table near IMMT1

Small Optics

- port with DC PD + RF PD + 2 x QPD needs at least periscope, beam shutter, 2 x lens, 3 x mirror, 3 x BS, 4 x beam dump

Wiki/SmallOptics

- there are

[before IMMT]

IMCREFL: DC PD + RF PD + 2 x QPD

IMCTrans: DC PD + 2 x QPD

REFL: DC PD + RF PD + 2 x QPD

ISS1: DC PD

ISS2: DC PD + 2 x QPD


[after IMMT]

POX: DC PD + RF PD ??

AS(SR2): DC PD + RF PD + 2 x QPD

ETMX: DC PD + 2 x QPD

ETMY: DC PD + 2 x QPD



LIGO-D1201103 (aLIGO ISCT1 table)

* one grid is 1inch x 1 inch

Some Important Info (for iKAGRA)

- Suspended optics from laser to PRM chamber are just IMC mirrors
 - we don't suspend IFI nor IMMTs
 - oplevs only for IMC mirrors (no oplevs for IMMTs)
- IMMTs are flat and have picomotors
- we don't need hardware interface between picomotor drivers and digital system

Questions

- Are we going to use IMMTs as actuators for ASC of FPMI?
- Are there PDs/QPDs which should be in vacuum?
 - only MIF REFL? **maybe none for iKAGRA**
- What are we going to do with MCo AR reflected beam?
 - I don't think it is essential for ISC point of view
Miyoki-san says it was in CLIO since IMC TRANS was low power
- How many standalone digital system will be needed by Nov 2014?
- What's the situation about FRC servo circuit? **Nakano has it**
 - Miyakawa-san said he delivered to IOO on Sep 2013
- Are we really going to use FRC? Is it a default plan?
- How many picomotors for a MC mirror? **2**
- What do we need for digital system to picomotor driver interfaces? Do we need some interface circuits? **no**
- What's the camera situation? How do we put them in the digital system?
- Where and how many do we have optical tables and pylons? What are the sizes of them? Do we need periscopes for them?